



CE DECLARATION OF CONFORMITY

Micromeritics Instrument Corporation, Inc.
4356 Communications Dr.
Norcross, Georgia 30093-2901 U.S.A.
ISO 9001:2000 Certified

VERIFICATION

Model Number: Gemini VII 2390
Type of Equipment: Physisorption Analyzer
Applicable Directive: 2004/108/EEC European Union (EU) Emissions and Immunity
73/23/EEC Low Voltage Directive European Union (EU) Safety

Standards of Conformity: EN 61326:1997+A1:1998+A2:2001
EN 61000-4-2:95 Electronic Discharge
EN 61000-4-3:96 RF Electromagnetic Field
EN 61000-4-4:95 Electrical Fast Transient/Burst
EN 61000-4-5:95 Surge
EN 61000-4-6:96 RF Common Mode
EN 61000-4-11:1994 Voltage Dips, Interruptions, and Variations
EN 61326 (CE) Radiated Emissions
EN 61326 (CE) Conducted Emissions
EN 61000-3-2:1995+A14:1998 Harmonic Current Emissions
EN 61000-3-3:1995 Voltage Fluctuations

EN 61010-1 Part 1: General Safety Requirements

European Contact Address: Micromeritics Germany GmbH, Avantis Science Park, Rutherford 108, D-52072 Aachen
Tel: +49 (0) 241 189 446 0, Fax: +49 (0) 241 189 446 11
email:micromeritics.de@micromeritics.com

Micromeritics Instrument Corporation hereby declares that the instrument(s) specified above conform(s) to the Directive(s) and Standard(s), when installed and operated in accordance with the specifications set forth by Micromeritics. The original of this document is kept at Micromeritics with copies of the relevant test data and certificates, which constitute the required technical file for self declaration.

Uth Khan
Signature

3-28-2011
Date

Stefan Koch
Print name

Director of Engineering
Position