



CE DECLARATION OF CONFORMITY

Micromeritics Instrument Corporation, Inc.
One Micromeritics Drive
Norcross, Georgia 30093-1877 U.S.A.
ISO 9001:2000 Certified

VERIFICATION

Model Number: Sedigraph 5120

Type of Equipment: Particle Size Analyzer

Directives: EMC Directive 89/336/EEC
Low Voltage Directive 73/23/EEC

Standards of Conformity: EN 61326: 1997 +A1: 1998
IEC 61000-4-2 Electronic Discharge
IEC 61000-4-3 RF Electromagnetic Field
IEC 61000-4-4 Electrical Fast Transient/Burst
IEC 61000-4-5 Surge
IEC 61000-4-6 RF Common Mode
IEC 61000-4-8 Power Frequency Magnetic Field
IEC 61000-4-11 Voltage Dips, Interruptions, and Variations

CISPR 11 Radiated Emissions
CISPR 11 Conducted Emissions
EN 61000-3-2 Harmonic Current Emissions
EN 61000-3-3 Voltage Fluctuations

EN 61010-1 Part 1: General Safety Requirements

Micromeritics Instrument Corporation hereby declares that the instrument(s) specified above conform(s) to the Directive(s) and Standard(s), when installed and operated in accordance with the specifications set forth by Micromeritics. The original of this document is kept at Micromeritics with copies of the relevant test data and certificates, which constitute the required technical file for self declaration.

Stefan Koch
Signature

10-18-2007
Date

Stefan Koch
Print name

Director of Engineering.
Position